

Off : 04862 224366

Fax: 04862 229586

# **AL-AZHAR DENTAL COLLEGE**

**(APPROVED BY DENTAL COUNCIL OF INDIA & AFFILIATED TO KUHS)**

**RUN BY NOORUL ISLAM TRUST**

**PERUMPILLICHIRA P. O., THODUPUZHA, IDUKKI DIST., KERALA- 685 605**

---

**ORAL MEDICINE AND RADIOLOGY**




**AL-AZHAR DENTAL COLLEGE**

**HOSPITAL INFECTION CONTROL COMMITTEE(HICC)**

**LOG BOOK -INFECTION CONTROL**

**ORAL MEDICINE AND RADIOLOGY**



**Prof. Dr. Harvey Thomas MDS**  
Principal  
Al-Azhar Dental College  
Thodupuzha -685 605

DEPARTMENT DMR DATE: 1/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
8 a.m	2:15 p.m		di

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
Bacillo	Yes		di

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
NO			di

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	Diagnostic instruments	Nexol	10 a.m		di
	"	"	11 a.m		di
	"	"	1 p.m		di

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	Diagnostic instruments.	10:00 a.m	10:45 a.m	45 minutes	121°C	25 psi	1/7/21 10:00 to 10:45		di
2	"	11:20 a.m	12:05 p.m	"	"	"	1/7/21 11:20 to 12:05		di
3	"	12:30 p.m	1:15 p.m	"	"	"	1/7/21 12:30 to 1:15		di
4	"	2:00 p.m	2:45 p.m	"	"	"	1/7/21 2:00 to 2:45		di
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT *DIAK*

DATE *2/7/21*

1. FLOOR CLEANING

TIME OF CLEANING	ROOM	HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC / MEMBER
<i>8 am</i>	<i>P.1 P.2 P.10</i>	<i>[Signature]</i>	<i>[Signature]</i>

2. CLEANING - DENTAL CHAIR

CHEMICAL DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC / MEMBER
<i>Bac 1101</i>	<i>Yes</i>	<i>[Signature]</i>	<i>[Signature]</i>

3. FUMIGATION

FUMIGATION CODE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<i>Yes</i>	<i>2:50 PM</i>	<i>[Signature]</i>	<i>[Signature]</i>

4. MANUAL CLEANING OF INSTRUMENTS

SL No.	SET OF INSTRUMENTS	CHEMICAL DISINFECTANT USED AND DURATION (YES/NO)	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE / HICC / MEMBER
	<i>Diagnostic Instruments</i>	<i>None</i>	<i>10 am</i>	<i>[Signature]</i>	<i>[Signature]</i>
	<i>"</i>	<i>None</i>	<i>11 am</i>	<i>[Signature]</i>	<i>[Signature]</i>
	<i>"</i>	<i>"</i>	<i>3 PM</i>	<i>[Signature]</i>	<i>[Signature]</i>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

SL No.	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC / MEMBER
		START	END	CYCLE LENGTH					
1	<i>Diagnostic Instruments</i>	<i>10:00 am</i>	<i>10:45 am</i>	<i>45 minutes</i>	<i>121°C</i>	<i>30 psi</i>	<i>2/7/21 10:00 to 10:45</i>	<i>[Signature]</i>	<i>[Signature]</i>
2	<i>"</i>	<i>11:20 am</i>	<i>12:05 pm</i>	<i>"</i>	<i>"</i>	<i>"</i>	<i>2/7/21 11:20 to 12:05</i>	<i>[Signature]</i>	<i>[Signature]</i>
3	<i>"</i>	<i>12:40 pm</i>	<i>1:15 pm</i>	<i>"</i>	<i>"</i>	<i>"</i>	<i>2/7/21 12:30 to 1:15</i>	<i>[Signature]</i>	<i>[Signature]</i>
4	<i>"</i>	<i>2:00 pm</i>	<i>2:45 pm</i>	<i>"</i>	<i>"</i>	<i>"</i>	<i>2/7/21 2:00 to 2:45</i>	<i>[Signature]</i>	<i>[Signature]</i>
5	<i>"</i>								
6									
7									

PREPARED AND APPROVED BY



*[Handwritten signature]*

Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DATE: 3/1/21

DEPARTMENT: DMK

1. FLOOR CLEANING

TIME OF CLEANING  
 MORNING: 8 am  
 NOON: 2:15 pm

HOUSEKEEPER/INCHARGE SIGNATURE: [Signature]

SIGNATURE OF THE STAFF INCHARGE / IICC MEMBER: [Signature]

2. CLEANING - DENTAL CHAIR

CHEMICAL/DISINFECTANT USED: Bacillol

CLEANED AFTER EACH PATIENT (YES/NO): Yes

ATTENDER INCHARGE (SIGNATURE): [Signature]

SIGNATURE OF THE STAFF INCHARGE / IICC MEMBER: [Signature]

3. FUMIGATION

FUMIGATION DONE YES/NO: No

TIME: [Blank]

SIGNATURE: [Signature]

REMARKS & SIGNATURE OF STAFF IN CHARGE: [Signature]

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL/DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE / IICC MEMBER
	Diagnostic Instruments	Nextal	10 am	[Signature]	[Signature]
	"	"	11 am	[Signature]	[Signature]
	"	"	1 pm	[Signature]	[Signature]

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl. No	SET OF INSTRUMENTS	TIME			TEMP °C	PRESSURE PSI	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / IICC MEMBER
		START	END	CYCLE LENGTH					
1	Diagnostic instrument	10:00 a.m.	10:45 a.m.	45 minutes	121°C	151	3/1/21 10:00 to 10:45	[Signature]	[Signature]
2	"	11:20 a.m.	12:05 p.m.	"	"	"	3/1/21 11:20 to 12:05	[Signature]	[Signature]
3	"	12:30 p.m.	1:15 p.m.	"	"	"	3/1/21 12:30 to 1:15	[Signature]	[Signature]
4	"	2:00 p.m.	2:45 p.m.	"	"	"	3/1/21 2:00 to 2:45	[Signature]	[Signature]
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS  
 Principal  
 Al-Azhar Dental College  
 Thodupuzha - 685 605

DEPARTMENT OMR

DATE: 5/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
8 a.m	2:15 p.m		di

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
Bacillo	Yes		di

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
YES	2:50 PM		di

4. MANUAL CLEANING OF INSTRUMENTS

SI. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	Diagnostic instruments	Hexol	10 a.m		di
	"	"	11 a.m		di
	"	"	21 p.m		di

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

SI.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	Diagnostic instruments	10:00 a.m	10:45 a.m	45 minutes	121°C	30 Psi	5/7/21 10:00 to 10:45		di
2	"	11:20 a.m	12:05 p.m	"	"	"	5/7/21 11:20 to 12:05		di
3	"	12:30 p.m	1:15 p.m	"	"	"	5/7/21 12:30 to 1:15		di
4	"	2:00 p.m	2:45 p.m	"	"	"	5/7/21 2:00 to 2:45		di
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT OMR

DATE: 6/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8 a.m</u>	<u>2:15 p.m</u>		<u>dci</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillo</u>	<u>Yes</u>		<u>dci</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>NO</u>			<u>dci</u>

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>Diagnostic Instruments</u>	<u>Hexitol</u>	<u>10 a.m</u>		<u>dci</u>
	<u>"</u>	<u>"</u>	<u>11 a.m</u>		<u>dci</u>
	<u>"</u>	<u>"</u>	<u>1 p.m</u>		<u>dci</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>Diagnostic instruments</u>	<u>10:00 a.m</u>	<u>10:45 a.m</u>	<u>45 minutes</u>	<u>121°C</u>	<u>15 psi</u>	<u>6/7/21 10:00 to 10:45</u>		<u>dci</u>
2	<u>"</u>	<u>11:20 a.m</u>	<u>12:05 p.m</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>6/7/21 11:20 to 12:05</u>		<u>dci</u>
3	<u>"</u>	<u>12:30 p.m</u>	<u>1:15 p.m</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>6/7/21 12:30 to 1:15</u>		<u>dci</u>
4	<u>"</u>	<u>2:00 p.m</u>	<u>2:45 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>6/7/21 2:00 to 2:45</u>		<u>dci</u>
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT P.M.E. DATE 4/2/22

### 1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
5 A.M	2:15 P.M	<i>[Signature]</i>	<i>[Signature]</i>

### 2. CLEANING - DENTAL CHAIR

CHEMICAL DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
<u>Briillol</u>	<u>Yes</u>	<i>[Signature]</i>	<i>[Signature]</i>

### 3. FUMIGATION

FUMIGATION DONE (YES/NO)	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>Yes</u>	<u>2:50 PM</u>	<i>[Signature]</i>	<i>[Signature]</i>

### 4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
	<u>Diagnostic instruments</u>	<u>Briillol</u>	<u>10:00 am</u>	<i>[Signature]</i>	<i>[Signature]</i>
	<u>"</u>	<u>"</u>	<u>11:00 am</u>	<i>[Signature]</i>	<i>[Signature]</i>
	<u>"</u>	<u>"</u>	<u>1 P.M</u>	<i>[Signature]</i>	<i>[Signature]</i>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
		START	END	CYCLE LENGTH					
1	<u>Diagnostic instruments</u>	<u>10:00 am</u>	<u>10:45 am</u>	<u>45 minutes</u>	<u>121°C</u>	<u>15 PS</u>	<i>[Indicator]</i>	<i>[Signature]</i>	<i>[Signature]</i>
2	<u>"</u>	<u>11:20 am</u>	<u>12:05 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<i>[Indicator]</i>	<i>[Signature]</i>	<i>[Signature]</i>
3	<u>"</u>	<u>12:30 pm</u>	<u>1:15 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<i>[Indicator]</i>	<i>[Signature]</i>	<i>[Signature]</i>
4	<u>"</u>	<u>2:00 pm</u>	<u>2:45 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<i>[Indicator]</i>	<i>[Signature]</i>	<i>[Signature]</i>
5									
6									
7									

PREPARED AND APPROVED BY




Prof. Dr. Harvey Thomas MD  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605




DEPARTMENT OMR DATE: 8/7/21 (23)


1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8 a.m</u>	<u>2:15 p.m</u>		<u>di</u>




2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Baillol</u>	<u>Yes</u>		<u>di</u>

3. FUMIGATION





FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>NO</u>			<u>di</u>

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>Diagnostic instruments</u>	<u>Hexitol</u>	<u>10 a.m</u>		<u>di</u>
	<u>"</u>	<u>"</u>	<u>11 a.m</u>		<u>di</u>
	<u>"</u>	<u>"</u>	<u>1 p.m</u>		<u>di</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	<u>Diagnostic instruments</u>	<u>10:00 a.m</u>	<u>10:45 a.m</u>	<u>45 minutes</u>	<u>121°</u>	<u>30 psi</u>	<u>8/7/21 10:00 to 10:45</u>		<u>di</u>
2	<u>"</u>	<u>11:20 a.m</u>	<u>12:05 p.m</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>8/7/21 11:20 to 12:05</u>		<u>di</u>
3	<u>"</u>	<u>12:30 p.m</u>	<u>1:15 PM</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>8/7/21 12:30 to 1:15</u>		<u>di</u>
4	<u>"</u>	<u>2:00 p.m</u>	<u>2:45 PM</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>8/7/21 2:00 to 2:45</u>		<u>di</u>
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT OMR

DATE 9/1/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8 a.m</u>	<u>2:15 p.m</u>		<u>di</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Enallo</u>	<u>Yes</u>		<u>di</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>Yes</u>	<u>9:50 PM</u>		<u>di</u>

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>Diagnostic instruments</u>	<u>Hexidal</u>	<u>10 a.m</u>		<u>di</u>
	<u>"</u>	<u>"</u>	<u>11 a.m</u>		<u>di</u>
	<u>"</u>	<u>"</u>	<u>1 p.m</u>		<u>di</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>Diagnostic instruments</u>	<u>10:00 a.m</u>	<u>10:45 a.m</u>	<u>45 minutes</u>	<u>121°C</u>	<u>30 psi</u>	<u>9/1/21 10:00 to 10:45</u>		<u>di</u>
2	<u>"</u>	<u>11:20 a.m</u>	<u>12:05 p.m</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>9/1/21 11:20 to 12:05</u>		<u>di</u>
3	<u>"</u>	<u>12:30 p.m</u>	<u>1:15 p.m</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>9/1/21 12:30 to 1:15</u>		<u>di</u>
4	<u>"</u>	<u>2:00 p.m</u>	<u>2:45 p.m</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>9/1/21 2:00 to 2:45</u>		<u>di</u>
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT OMR

DATE: 10/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8 a.m</u>	<u>2:15 p.m</u>		<u>di</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
<u>Bactinol</u>	<u>Yes</u>		<u>di</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>NO</u>			<u>di</u>

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>Diagnostic instruments</u>	<u>Hexidol</u>	<u>10:00 a.m</u>		<u>di</u>
	<u>)</u>	<u>)</u>	<u>11:00 a.m</u>		<u>di</u>
	<u>)</u>	<u>)</u>	<u>1:00 p.m</u>		<u>di</u>

PREPARED AND APPROVED BY

STERILIZATION LOG SHEET

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	<u>Diagnostic instruments.</u>	<u>10:00 a.m</u>	<u>10:45 p.m</u>	<u>45 minutes</u>	<u>1212</u>	<u>30 psi</u>	<u>10/7/21 10:00 to 10:45</u>		<u>di</u>
2	<u>)</u>	<u>11:20 a.m</u>	<u>12:05 p.m</u>	<u>)</u>	<u>)</u>	<u>)</u>	<u>10/7/21 11:20 to 12:05</u>		<u>di</u>
3	<u>)</u>	<u>12:30 p.m</u>	<u>1:15 p.m</u>	<u>)</u>	<u>)</u>	<u>)</u>	<u>10/7/21 12:30 to 1:15</u>		<u>di</u>
4	<u>)</u>	<u>2:00 p.m</u>	<u>2:45 p.m</u>	<u>)</u>	<u>)</u>	<u>)</u>	<u>10/7/21 2:00 to 2:45</u>		<u>di</u>
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT OMF DATE: 12/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8:00</u>	<u>2:15 PM</u>		<u>di</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillo 1</u>	<u>Yes.</u>		<u>di</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>Yes</u>	<u>2:50 PM</u>		<u>di</u>

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>Diagnostic Instruments</u>	<u>Hexsol</u>	<u>10:00 am</u>		<u>di</u>
	<u>"</u>	<u>"</u>	<u>11:00 am</u>		<u>di</u>
	<u>"</u>	<u>"</u>	<u>1:00 pm</u>		<u>di</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>Diagnostic instruments.</u>	<u>10:00 am</u>	<u>10:45 am</u>	<u>45 mins</u>	<u>121°C</u>	<u>30 Psi</u>	<u>12/7/21 10:00 to 10:45</u>		<u>di</u>
2	<u>"</u>	<u>11:20 am</u>	<u>12:05 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>12/7/21 11:20 to 12:05</u>		<u>di</u>
3	<u>"</u>	<u>12:30 pm</u>	<u>1:15 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>12/7/21 12:30 to 1:15</u>		<u>di</u>
4	<u>"</u>	<u>2:00 pm</u>	<u>2:45 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>12/7/21 2:00 to 2:45</u>		<u>di</u>
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT DMR

DATE: 13/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		di
8:00	2:15pm		

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
bac1101	yes		di

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
NO			di

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	Diagnostic instrument	Henzol	10:00 am		di
	"	"	11:00 am		di
	"	"	1:00 pm		di


PREPARED AND APPROVED BY

STAFF INCHARGE FOR SHEET

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/HICC MEMBER.
		START	END	CYCLE LENGTH					
1	Diagnostic Instrument	10:00 am	10:45 am	45 min	121°C	30 Psi	13/7/21 10:00 to 10:45		di
2	"	11:30 am	12:05 pm	"	"	"	13/7/21 11:20 to 12:05		di
3	"	12:30 pm	1:15 pm	"	"	"	13/7/21 12:30 to 1:15		di
4	"	2:00 pm	2:45 pm	"	"	"	13/7/21 2:00 to 2:45		di
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha-685 605

DEPARTMENT OMR

DATE: 14/7/21

125

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8 am</u>	<u>2:15 pm</u>	<u>[Signature]</u>	<u>di</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillo 1</u>	<u>Yes</u>	<u>[Signature]</u>	<u>di</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>YES</u>	<u>2:50 PM</u>	<u>[Signature]</u>	<u>di</u>

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>diagnostic instruments</u>	<u>HexiJo 1</u>	<u>10 am</u>	<u>[Signature]</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>11 am</u>	<u>[Signature]</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>1 pm</u>	<u>[Signature]</u>	<u>di</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>Diagnostic Instruments</u>	<u>10:00 am</u>	<u>10:45 am</u>	<u>45 min</u>	<u>121°C</u>	<u>30 Psi</u>	<u>14/7/21 10:00 to 10:45</u>	<u>[Signature]</u>	<u>di</u>
2	<u>"</u>	<u>11:20 am</u>	<u>12:05 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>14/7/21 11:20 to 12:05</u>	<u>[Signature]</u>	<u>di</u>
3	<u>"</u>	<u>12:30 pm</u>	<u>1:15 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>14/7/21 12:30 to 1:15</u>	<u>[Signature]</u>	<u>di</u>
4	<u>"</u>	<u>2:00 pm</u>	<u>2:45 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>14/7/21 2:00 to 2:45</u>	<u>[Signature]</u>	<u>di</u>
5									
6									
7									

PREPARED AND APPROVED BY



[Handwritten Signature]

Prof. Dr. Harvey Thomas MI  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT OMR

DATE: 15/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8 am</u>	<u>2:15 pm</u>	<u>M</u>	<u>di</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillo 1</u>	<u>Yes</u>	<u>R</u>	<u>di</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>NO</u>		<u>R</u>	<u>di</u>

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>diagnostic instrument</u>	<u>Hexitol</u>	<u>10 am</u>	<u>R</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>11 am</u>	<u>R</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>1 pm</u>	<u>R</u>	<u>di</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>diagnostic instrument</u>	<u>10:00 am</u>	<u>10:45 am</u>	<u>45 min</u>	<u>121°C</u>	<u>30 PSI</u>	<u>15/7/21 10:00 to 10:45</u>	<u>R</u>	<u>di</u>
2	<u>"</u>	<u>11:20 am</u>	<u>12:05 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>15/7/21 11:20 to 12:05</u>	<u>R</u>	<u>di</u>
3	<u>"</u>	<u>1:30 pm</u>	<u>1:15 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>15/7/21 12:30 to 1:15</u>	<u>R</u>	<u>di</u>
4	<u>"</u>	<u>2:00 pm</u>	<u>2:45 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>15/7/21 2:00 to 2:45</u>	<u>R</u>	<u>di</u>
5									
6									
7									

PREPARED AND APPROVED BY



*[Handwritten signature]*

DEPARTMENT OMR

DATE: 16/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8am</u>	<u>2:15pm</u>	<u>[Signature]</u>	<u>[Signature]</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillo1</u>	<u>Yes</u>	<u>[Signature]</u>	<u>[Signature]</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>YES</u>	<u>2:50 PM</u>	<u>[Signature]</u>	<u>[Signature]</u>

4. MANUAL CLEANING OF INSTRUMENTS

SI. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>diagnostic instrument</u>	<u>Hexital</u>	<u>10am</u>	<u>[Signature]</u>	<u>[Signature]</u>
	<u>"</u>	<u>"</u>	<u>11am</u>	<u>[Signature]</u>	<u>[Signature]</u>
	<u>"</u>	<u>"</u>	<u>1pm</u>	<u>[Signature]</u>	<u>[Signature]</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

SI.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/HICC MEMBER
		START	END	CYCLE LENGTH					
1	<u>diagnostic instrument</u>	<u>10:00 am</u>	<u>10:45 am</u>	<u>45 min</u>	<u>171c</u>	<u>30 Psi</u>	<u>16/7/21 10:00 to 10:45</u>	<u>[Signature]</u>	<u>[Signature]</u>
2	<u>"</u>	<u>11:30 am</u>	<u>12:05 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>16/7/21 11:30 to 12:05</u>	<u>[Signature]</u>	<u>[Signature]</u>
3	<u>"</u>	<u>12:30 pm</u>	<u>1:15 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>16/7/21 12:30 to 1:15</u>	<u>[Signature]</u>	<u>[Signature]</u>
4	<u>"</u>	<u>2:00pm</u>	<u>2:45pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>16/7/21 2:00 to 2:45</u>	<u>[Signature]</u>	<u>[Signature]</u>
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605



DEPARTMENT OMK

DATE: 17/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8am</u>	<u>2:15pm</u>	<u>Pz</u>	<u>di</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillo 1</u>	<u>Yes</u>	<u>Pz</u>	<u>di</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>NO</u>		<u>Pz</u>	<u>di</u>

4. MANUAL CLEANING OF INSTRUMENTS


Sl.No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>Diagnostic Instrument</u>	<u>Hexits 1</u>	<u>10am</u>	<u>Pz</u>	<u>di</u>
	<u>"</u>	<u>J3</u>	<u>11am</u>	<u>Pz</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>1pm</u>	<u>Pz</u>	<u>di</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>Diagnostic Instrument</u>	<u>10:00 am</u>	<u>10:45 am</u>	<u>45 min</u>	<u>121°C</u>	<u>30 Psi</u>	<u>17/7/21 10:00 to 10:45</u>	<u>Pz</u>	<u>di</u>
2	<u>"</u>	<u>11:20 am</u>	<u>12:05 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>17/7/21 11:20 to 12:05</u>	<u>Pz</u>	<u>di</u>
3	<u>"</u>	<u>12:30 pm</u>	<u>1:15 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>17/7/21 12:30 to 1:15</u>	<u>Pz</u>	<u>di</u>
4	<u>"</u>	<u>2:00 pm</u>	<u>2:45 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>17/7/21 2:00 to 2:45</u>	<u>Pz</u>	<u>di</u>
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT OMR DATE: 19/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8 am</u>	<u>2:15 pm</u>	<u>P</u>	<u>di</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillo 1</u>	<u>Yes</u>	<u>P</u>	<u>di</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>YES</u>	<u>2:50 PM</u>	<u>P</u>	<u>di</u>

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>diagnostic instrument</u>	<u>Hexitol</u>	<u>10 am</u>	<u>P</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>11 am</u>	<u>P</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>1 pm</u>	<u>P</u>	<u>di</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>diagnostic instrument</u>	<u>10:00 am</u>	<u>10:45 am</u>	<u>45 min</u>	<u>121</u>	<u>30PSI</u>	<u>19/7/21 10:00 to 10:45</u>	<u>P</u>	<u>di</u>
2	<u>"</u>	<u>11:00 am</u>	<u>12:05 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>19/7/21 11:00 to 12:05</u>	<u>P</u>	<u>di</u>
3	<u>"</u>	<u>12:30 pm</u>	<u>1:15 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>19/7/21 12:30 to 1:15</u>	<u>P</u>	<u>di</u>
4	<u>"</u>	<u>2:00 pm</u>	<u>2:45 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>19/7/21 2:00 to 2:45</u>	<u>P</u>	<u>di</u>
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT OMR DATE: 20/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8am</u>	<u>2.15pm</u>	<u>PW</u>	<u>di</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillol</u>	<u>Yes</u>	<u>PW</u>	<u>di</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>No</u>		<u>PW</u>	<u>di</u>

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>diagnostic instrument</u>	<u>Hexachol</u>	<u>10am</u>	<u>PW</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>11am</u>	<u>PW</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>1pm</u>	<u>PW</u>	<u>di</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>diagnostic instrument</u>	<u>10:00 am</u>	<u>10:45 am</u>	<u>45min</u>	<u>121°C</u>	<u>30 Psi</u>	<u>20/7/21 10:00 to 10:45</u>	<u>PW</u>	<u>di</u>
2	<u>"</u>	<u>11:00 am</u>	<u>12:05 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>20/7/21 11:20 to 12:05</u>	<u>PW</u>	<u>di</u>
3	<u>"</u>	<u>12:30 pm</u>	<u>1:15 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>20/7/21 12:30 to 1:15</u>	<u>PW</u>	<u>di</u>
4	<u>"</u>	<u>2:00 pm</u>	<u>2:45 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>20/7/21 2:00 to 2:45</u>	<u>PW</u>	<u>di</u>
5									
6									
7									

PREPARED AND APPROVED BY



*(Handwritten signature)*

Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT OMR

DATE: 22/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8:30am</u>	<u>2:15pm</u>	<u>PV</u>	<u>di</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Excilol</u>	<u>Yes</u>	<u>PV</u>	<u>di</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>YES</u>	<u>2:50 PM</u>	<u>PV</u>	<u>di</u>

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDEE INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>Diagnostic instrument</u>	<u>Hexitol</u>	<u>10am</u>	<u>PV</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>11am</u>	<u>PV</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>1pm</u>	<u>PV</u>	<u>di</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>Diagnostic instrument</u>	<u>10:am</u>	<u>10:45am</u>	<u>45min</u>	<u>121°C</u>	<u>30psi</u>	<u>22/7/21 10:00 to 10:45</u>	<u>PV</u>	<u>di</u>
2	<u>"</u>	<u>11:20am</u>	<u>12:05pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>22/7/21 11:20 to 12:05</u>	<u>PV</u>	<u>di</u>
3	<u>"</u>	<u>12:30pm</u>	<u>1:15pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>22/7/21 12:30 to 1:15</u>	<u>PV</u>	<u>di</u>
4	<u>"</u>	<u>2:00pm</u>	<u>2:45pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>22/7/21 2:00 to 2:45</u>	<u>PV</u>	<u>di</u>
5									
6									
7									

PREPARED AND APPROVED BY



*[Handwritten signature]*

Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT OMR

DATE: 23/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8am</u>	<u>2.15pm</u>	<u>Pr</u>	<u>di</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
<u>Bacillol</u>	<u>Yes</u>	<u>Pr</u>	<u>di</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>No</u>		<u>Pr</u>	<u>di</u>

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>Diagnostic instrument</u>	<u>Hercol</u>	<u>10am</u>	<u>Pr</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>11am</u>	<u>Pr</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>1pm</u>	<u>Pr</u>	<u>di</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	<u>Diagnostic instrument</u>	<u>10am</u>	<u>10.45pm</u>	<u>25min</u>	<u>121°C</u>	<u>30psi</u>	<u>23/7/21 10-00 to 10-45</u>	<u>Pr</u>	<u>di</u>
2	<u>"</u>	<u>11-20am</u>	<u>12:05pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>23/7/21 11-20 to 12-05</u>	<u>Pr</u>	<u>di</u>
3	<u>"</u>	<u>12-30pm</u>	<u>1.15pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>23/7/21 12-30 to 1-15</u>	<u>Pr</u>	<u>di</u>
4	<u>"</u>	<u>2.00pm</u>	<u>2.45pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>23/7/21 2-00 to 2-45</u>	<u>Pr</u>	<u>di</u>
5									
6									
7									

PREPARED AND APPROVED BY



*Handwritten signature in blue ink.*

Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT OMR

DATE: 24/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>Done</u>	<u>2.15 pm</u>	<u>PV</u>	<u>di</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillol</u>	<u>Yes</u>	<u>PV</u>	<u>di</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>YES</u>	<u>2.50 PM</u>	<u>PV</u>	<u>di</u>

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>Diagnostic instrument</u>	<u>Hexibel</u>	<u>10am</u>	<u>PV</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>11am</u>	<u>PV</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>1pm</u>	<u>PV</u>	<u>di</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	<u>Diagnostic instrument</u>	<u>10am</u>	<u>10.45am</u>	<u>45min</u>	<u>121C</u>	<u>30Psi</u>	<u>24/7/21 10.00 to 10.45</u>	<u>PV</u>	<u>di</u>
2	<u>"</u>	<u>11.20am</u>	<u>12.05pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>24/7/21 11.20 to 12.05</u>	<u>PV</u>	<u>di</u>
3	<u>"</u>	<u>12.30pm</u>	<u>1.15pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>24/7/21 12.30 to 1.15</u>	<u>PV</u>	<u>di</u>
4	<u>"</u>	<u>2.00pm</u>	<u>2.45pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>24/7/21 2.00 to 2.45</u>	<u>PV</u>	<u>di</u>
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT OMR DATE: 26/7/21 134

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
8 am	2.15 pm	P	di

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
Banitol	Yes	P	di

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
NO		P	di

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	Diagnostic instrument	Hexidol	10 am	P	di
	"	"	11 am	P	di
	"	"	1 pm	P	di

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	Diagnostic instrument	10 am	10.45 pm	45 min	121°C	30 Psi	26/7/21 10:00 to 10:45	P	di
2	"	11.20 pm	12.05 pm	"	"	"	26/7/21 11:20 to 12:05	P	di
3	"	12.30 pm	1.15 pm	"	"	"	26/7/21 12:30 to 1:15	P	di
4	"	2.00 pm	2.45 pm	"	"	"	26/7/21 2:00 to 2:45	P	di
5									
6									
7									

PREPARED AND APPROVED BY



*[Handwritten signature]*

Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT OMR

DATE: 27/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8 am</u>	<u>2.15 pm</u>	<i>[Signature]</i>	<i>[Signature]</i>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Baukol</u>	<u>YES</u>	<i>[Signature]</i>	<i>[Signature]</i>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>YES</u>	<u>2.50 PM</u>	<i>[Signature]</i>	<i>[Signature]</i>

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>Diagnostic instrument</u>	<u>Hexitol</u>	<u>10 am</u>	<i>[Signature]</i>	<i>[Signature]</i>
	"	"	<u>11 am</u>	<i>[Signature]</i>	<i>[Signature]</i>
	"	"	<u>1 pm</u>	<i>[Signature]</i>	<i>[Signature]</i>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>Diagnostic instrument</u>	<u>10 am</u>	<u>10.45 pm</u>	<u>45 min</u>	<u>17°C</u>	<u>30 Psi</u>	<u>27/7/21</u> <u>10:00 to</u> <u>10:45</u>	<i>[Signature]</i>	<i>[Signature]</i>
2	"	<u>11.20 pm</u>	<u>12.05 pm</u>	"	"	"	<u>27/7/21</u> <u>11:20 to</u> <u>12:05</u>	<i>[Signature]</i>	<i>[Signature]</i>
3	"	<u>12.30 pm</u>	<u>1.15 pm</u>	<u>45</u>	"	"	<u>27/7/21</u> <u>12:30 to</u> <u>1:15</u>	<i>[Signature]</i>	<i>[Signature]</i>
4	"	<u>2.00 pm</u>	<u>2.45 pm</u>	"	"	"	<u>27/7/21</u> <u>2:00 to</u> <u>2:45</u>	<i>[Signature]</i>	<i>[Signature]</i>
5									
6									
7									

PREPARED AND APPROVED BY



*[Handwritten Signature]*

Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605



DEPARTMENT OMR

DATE: 28/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8am</u>	<u>2.15pm</u>	<u>[Signature]</u>	<u>[Signature]</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillo</u>	<u>YES</u>	<u>[Signature]</u>	<u>[Signature]</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>NO</u>	<u>1/8 D.</u>	<u>[Signature]</u>	<u>[Signature]</u>

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>diagnostic instrument</u>	<u>Hexitol</u>	<u>10am</u>	<u>[Signature]</u>	<u>[Signature]</u>
	<u>"</u>	<u>"</u>	<u>11am</u>	<u>[Signature]</u>	<u>[Signature]</u>
	<u>"</u>	<u>"</u>	<u>1pm</u>	<u>[Signature]</u>	<u>[Signature]</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	<u>diagnostic instrument</u>	<u>10am</u>	<u>10.15am</u>	<u>45min</u>	<u>121°</u>	<u>30Psi</u>	<u>28/7/21 10:00 to 10:45</u>	<u>[Signature]</u>	<u>[Signature]</u>
2	<u>"</u>	<u>11:20pm</u>	<u>12:05pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>28/7/21 11:20 to 12:05</u>	<u>[Signature]</u>	<u>[Signature]</u>
3	<u>"</u>	<u>12:30pm</u>	<u>1:15pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>28/7/21 12:30 to 1:15</u>	<u>[Signature]</u>	<u>[Signature]</u>
4	<u>"</u>	<u>2:06pm</u>	<u>2:45pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>28/7/21 2:00 to 2:45</u>	<u>[Signature]</u>	<u>[Signature]</u>
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT OMR

DATE: 29/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8 am</u>	<u>2:15 pm</u>	<u>[Signature]</u>	<u>[Signature]</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillool</u>	<u>YES</u>	<u>[Signature]</u>	<u>[Signature]</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>YES</u>	<u>2:50 PM</u>	<u>[Signature]</u>	<u>[Signature]</u>

4. MANUAL CLEANING OF INSTRUMENTS

Sr No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>diagnostic instrument</u>	<u>Hex2fol</u>	<u>10 am</u>	<u>[Signature]</u>	<u>[Signature]</u>
	<u>"</u>	<u>"</u>	<u>11 am</u>	<u>[Signature]</u>	<u>[Signature]</u>
	<u>"</u>	<u>"</u>	<u>1 pm</u>	<u>[Signature]</u>	<u>[Signature]</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

S/No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>diagnostic instrument</u>	<u>10 am</u>	<u>10:45 am</u>	<u>45 min</u>	<u>121°C</u>	<u>30PSI</u>	<u>29/7/21 10:00 to 10:45</u>	<u>[Signature]</u>	<u>[Signature]</u>
2	<u>"</u>	<u>1:20 pm</u>	<u>1:05 pm</u>	<u>1"</u>	<u>"</u>	<u>"</u>	<u>29/7/21 11:20 to 12:05</u>	<u>[Signature]</u>	<u>[Signature]</u>
3	<u>"</u>	<u>12:30 pm</u>	<u>1:15 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>29/7/21 12:30 to 1:15</u>	<u>[Signature]</u>	<u>[Signature]</u>
4	<u>"</u>	<u>2:00 pm</u>	<u>2:45 pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>29/7/21 2:00 to 2:45</u>	<u>[Signature]</u>	<u>[Signature]</u>
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT: ORL DATE: 30/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8am</u>	<u>2:15 pm</u>	<u>[Signature]</u>	<u>di</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillo</u>	<u>YES</u>	<u>[Signature]</u>	<u>di</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>No</u>	<u>Yes</u>	<u>[Signature]</u>	<u>di</u>

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>diagnostic instrument</u>	<u>HexP-Eol</u>	<u>10am</u>	<u>[Signature]</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>11am</u>	<u>[Signature]</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>1pm</u>	<u>[Signature]</u>	<u>di</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>diagnostic instruments</u>	<u>10am</u>	<u>10:45</u>	<u>45min</u>	<u>121°</u>	<u>30psi</u>	<u>30/7/21 10:00 to 10:45</u>	<u>[Signature]</u>	<u>di</u>
2	<u>"</u>	<u>11:20pm</u>	<u>12:05pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>30/7/21 11:20 to 12:05</u>	<u>[Signature]</u>	<u>di</u>
3	<u>"</u>	<u>0:30pm</u>	<u>1:15pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>30/7/21 12:30 to 1:15</u>	<u>[Signature]</u>	<u>di</u>
4	<u>"</u>	<u>2:00pm</u>	<u>2:45pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>30/7/21 2:00 to 2:45</u>	<u>[Signature]</u>	<u>di</u>
5									
6									
7									

PREPARED AND APPROVED BY



[Signature]

Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605

DEPARTMENT OMR DATE: 31/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8am</u>	<u>2.15pm</u>	<u>[Signature]</u>	<u>di</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillol</u>	<u>YES</u>	<u>[Signature]</u>	<u>di</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>YES</u>	<u>2.50PM</u>	<u>[Signature]</u>	<u>di</u>

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>diagnostic instrument</u>	<u>Hexel</u>	<u>10am</u>	<u>[Signature]</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>11am</u>	<u>[Signature]</u>	<u>di</u>
	<u>"</u>	<u>"</u>	<u>1pm</u>	<u>[Signature]</u>	<u>di</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	<u>diagnostic instruments</u>	<u>10am</u>	<u>10.45</u>	<u>45min</u>	<u>171°C</u>	<u>35PSI</u>	<u>31/7/21 10:50 to 10:45</u>	<u>[Signature]</u>	<u>di</u>
2	<u>"</u>	<u>11.20pm</u>	<u>12.05pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>31/7/21 11:20 to 12:05</u>	<u>[Signature]</u>	<u>di</u>
3	<u>"</u>	<u>12.30pm</u>	<u>1.15pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>31/7/21 12:30 to 1:15</u>	<u>[Signature]</u>	<u>di</u>
4	<u>"</u>	<u>2.00pm</u>	<u>2.45pm</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>31/7/21 2:00 to 2:45</u>	<u>[Signature]</u>	<u>di</u>
5									
6									
7									

PREPARED AND APPROVED BY



[Signature]

Prof. Dr. Harvey Thomas MDS  
Principal  
Al-Azhar Dental College  
Thodupuzha - 685 605