

Off : 04862 224366

Fax: 04862 229586


AL-AZHAR DENTAL COLLEGE

(APPROVED BY DENTAL COUNCIL OF INDIA & AFFILIATED TO KUHS)

RUN BY NOORUL ISLAM TRUST

PERUMPILLICHIRA P. O., THODUPUZHA, IDUKKI DIST., KERALA- 685 605

ORTHODONTICS



AL-AZHAR DENTAL COLLEGE
HOSPITAL INFECTION CONTROL COMMITTEE(HICC)
LOG BOOK -INFECTION CONTROL
ORTHODONTICS



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha -685 605

DEPARTMENT Orthodontics

DATE: 1/7/21

115

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
8:20	2:30	<i>Sabene</i>	<i>BT</i>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<i>B-cillox</i>	<i>Yes</i>	<i>[Signature]</i>	<i>BT</i>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<i>Y</i>			

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<i>OP instruments</i>	<i>Cidex Sterilade</i>	<i>8:00am / 1:45pm</i>	<i>[Signature]</i>	<i>BT</i>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	<i>Diagnostic tray</i>	<i>8:45</i>	<i>9:00</i>	<i>45 min</i>	<i>121C</i>	<i>30psi</i>	<i>1/7/21 8:55am to 9:00am</i>	<i>[Signature]</i>	<i>BT</i>
2	<i>Instrument tray</i>	"	"	"	"	"			
3	<i>Orthodontic pliers</i>	"	"	"	"	"	<i>1/7/21 8:55am to 9:00am</i>		
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT: Orthodontics DATE: 2/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
8:00	2:00	<i>Sabeena</i>	<i>BT</i>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<i>Bacillus</i>	<i>yes</i>	<i>[Signature]</i>	<i>BT</i>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL/DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<i>op 7 instruments</i>	<i>cidex glutaraldehyde</i>	<i>8:00am / 1:45pm</i>	<i>[Signature]</i>	<i>BT</i>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No.	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	<i>orthodontic instruments</i>	<i>8:45</i>	<i>9:00</i>	<i>15 min</i>	<i>121</i>	<i>30 psi</i>	<i>2/7/21 8:15am to 9:00am</i>	<i>[Signature]</i>	<i>BT</i>
2	<i>Impression trays</i>	<i>9</i>	<i>11</i>	<i>11</i>	<i>121</i>	<i>11</i>		<i>[Signature]</i>	<i>[Signature]</i>
3	<i>orthodontic pliers</i>	<i>9</i>	<i>9</i>	<i>9</i>	<i>121</i>	<i>11</i>	<i>2/7/21 8:15am to 9:00am</i>	<i>[Signature]</i>	<i>[Signature]</i>
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Orthodontics

DATE: 3/7/21

117

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
8:00	2:00	Sabeena	BT

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
1% methyl	Yes	[Signature]	BT

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	op instruments	Cidex glutaraldehyde	8:00am / 1:45 pm	[Signature]	BT

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No.	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	diagnostic instruments	8:15	9:00	45min	121C 30psi	30 psi	3/7/21 8:15am to 9:00am	[Signature]	[Signature]
2	Impression trays	"	"	"	"	"	3/7/21 8:15am to 9:00am	[Signature]	[Signature]
3	orthodontic pliers	"	"	"	"	"			
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Orthodontics

DATE: 5/7/21

118

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
8:00	2:00	<u>Sabeeka</u>	<u>BT</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>B actol</u>	<u>yes</u>	<u>[Signature]</u>	<u>BT</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>yes</u>	<u>3:00 pm.</u>	<u>[Signature]</u>	<u>[Signature]</u>

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
	<u>Op instruments</u>	<u>cinex 5 minutes</u>	<u>8:00am / 1:45pm</u>	<u>[Signature]</u>	<u>BT</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
		START	END	CYCLE LENGTH					
1	<u>diagnostic instruments</u>	<u>8:15</u>	<u>9:00</u>	<u>45min</u>	<u>121c</u>	<u>30psi</u>	<u>5/7/21 8:15am to 9:00am</u>	<u>[Signature]</u>	<u>[Signature]</u>
2	<u>impression tray</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>		<u>[Signature]</u>	<u>[Signature]</u>
3	<u>orthodontic pliers</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>5/7/21 8:15am to 9:00am</u>	<u>[Signature]</u>	<u>[Signature]</u>
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Orthodontics

DATE: 6/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
8:00	2:00	<u>Sabeena</u>	<u>BT</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bouillol</u>	<u>Yes</u>	<u>[Signature]</u>	<u>BT</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>up instruments</u>	<u>Cidex</u> <u>metronaldehyde</u>	<u>8:00am</u> <u>1:45pm</u>	<u>[Signature]</u>	<u>BT</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	<u>Diagnostic instruments</u>	<u>8:15</u>	<u>9:00</u>	<u>45min</u>	<u>121c</u>	<u>30psi</u>	<u>6/7/21</u> <u>8:15am to</u> <u>9:00am</u>	<u>[Signature]</u>	<u>BT</u>
2	<u>Impression tray</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>			
3	<u>orthodontic pliers</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>			
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT: Orthodontic

DATE: 7/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
8:00	2:00	<i>Subena</i>	<i>BT</i>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
Bacillox	Yes	<i>[Signature]</i>	<i>BT</i>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS / SIGNATURE OF STAFF IN CHARGE
Yes	3:00pm	<i>[Signature]</i>	<i>BT</i>

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	op instruments	cidex 1:45 min	8:00am / 1:45pm	<i>[Signature]</i>	<i>BT</i>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	Diagnostic Instruments	8:15	9:00	45min	121c	30psi	7/7/21 8:15am to 9:00am	<i>[Signature]</i>	<i>BT</i>
2	Impression trays	"	"	"	"	"			
3	orthodontic pliers	"	"	"	"	"	7/7/21 8:15am to 9:00am		
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha-685 605

DEPARTMENT

Orthodontics

DATE: 8/7/21

121

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
8:00	2:00	Sarben	BJ

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
B cotton	Yes	[Signature]	BJ

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	op instruments	cidex 20 minutes	8:00am/1:45pm	A	BJ

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	Diagno she instruments	8:15	9:00	15 min	121C	30PSI	8/7/21 8:15am to 9:00am	[Signature]	[Signature]
2	Impressors tray	"	"	"	"	"			
3	Orthodontic pliers	"	"	"	"	"	8/7/21 8:15am to 9:00am		
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Oral-dental

DATE: 9/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
8:00	2:00	Sabeena	BT

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
Betadine	Yes	[Signature]	BT

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
Yes	3:00pm	[Signature]	BT

4. MANUAL CLEANING OF INSTRUMENTS


SI. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	OP instruments	Cidex extra	8:00am / 1:45pm	[Signature]	BT

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

SI.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	Diagnostic Instruments	8:15	9:00	45min	121°C	30 psi	9/7/21 8:15am to 9:00am	[Signature]	BT
2	Impression tray	"	"	"	"	"			
3	Orthodontic pliers	"	"	"	"	"	9/7/21 8:15am to 9:00am		
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT

Orthodontics

DATE:

10/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<i>8:00 am.</i>	<i>2:00pm</i>	<i>Sabena</i>	<i>BT</i>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<i>Bacillo.</i>	<i>Yes</i>	<i>L</i>	<i>BT</i>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<i>OP Instruments</i>	<i>CIDCP Glutaraldehyde</i>	<i>8:00am/1:45pm</i>		<i>BT</i>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	<i>Diagnostic Instruments</i>	<i>8:15am</i>	<i>9:00am</i>	<i>45min</i>	<i>121°C</i>	<i>30psi</i>	<i>10/7/21 8:15am to 9:00am</i>	<i>[Signature]</i>	<i>BT</i>
2	<i>Impression Trays</i>	<i>"</i>	<i>"</i>	<i>"</i>	<i>"</i>	<i>"</i>		<i>[Signature]</i>	
3	<i>Orthodontic pliers</i>	<i>"</i>	<i>"</i>	<i>"</i>	<i>"</i>	<i>"</i>	<i>10/7/21 8:15am to 9:00am</i>	<i>[Signature]</i>	
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Orthodontics

DATE: 12/7/21 124

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8:00am</u>	<u>01:00pm</u>	<u>Sabeena</u>	<u>B.T.</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillol.</u>	<u>Yes</u>	<u>[Signature]</u>	<u>B.T.</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>Yes</u>	<u>3:00pm.</u>	<u>[Signature]</u>	<u>B.T.</u>

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>OP Instruments</u>	<u>Cloc-p glutaraldehyde</u>	<u>8:00am / 1:45pm</u>	<u>[Signature]</u>	<u>B.T.</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/ F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>Diagnostic Instruments</u>	<u>8:15am</u>	<u>9:00am</u>	<u>45min</u>	<u>121°C</u>	<u>30psi</u>	<u>12/7/21 8:15am to 9:00am</u>	<u>[Signature]</u>	<u>B.T.</u>
2	<u>Impression Trays</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>			
3	<u>Orthodontic pliers</u>	<u>"</u>	<u>"</u>	<u>7</u>	<u>"</u>	<u>"</u>	<u>12/7/21 8:15am to 9:00am</u>		
4									
5									
6									
7									



PREPARED AND APPROVED BY

Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT: Orthodontics DATE: 13/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
8:00am	2:00pm	<u>Sabena</u>	<u>B.T.</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL/DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacitob.</u>	<u>Yes</u>	<u>[Signature]</u>	<u>B.T.</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL/DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>Op Instruments</u>	<u>CIDEX glutaraldehyde</u>	<u>8:00am/11:45pm</u>	<u>[Signature]</u>	<u>B.T.</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>Diagnostic tray</u>	<u>8:15am</u>	<u>9:00am</u>	<u>45min</u>	<u>121°C</u>	<u>30psi</u>	<u>[Stamp: 13/7/21 8:15am to 9:00am]</u>	<u>[Signature]</u>	
2	<u>Impression tray</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>		<u>[Signature]</u>	<u>B.T.</u>
3	<u>Orthodontic wax</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>[Stamp: 13/7/21 8:15am to 9:00am]</u>	<u>[Signature]</u>	
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Orthodontics

DATE: 14/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8:00AM</u>	<u>2:00PM</u>	<u>Sabreem</u>	<u>B.T.</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bocillo7</u>	<u>Yes</u>	<u>[Signature]</u>	<u>B.T.</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>Yes</u>	<u>3:00pm</u>	<u>[Signature]</u>	<u>B.T.</u>

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>O.P instruments</u>	<u>CIDEX glutaraldehyde</u>	<u>8:00am / 1:45pm</u>	<u>[Signature]</u>	<u>B.T.</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	<u>Diagnosis tray</u>	<u>8:15am</u>	<u>9:00am</u>	<u>45min</u>	<u>121°C</u>	<u>30 psi</u>	<u>14/7/21 8:15am to 9:00am</u>	<u>[Signature]</u>	<u>B.T.</u>
2	<u>Impression tray</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>			
3	<u>Orthodontic pliers</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>14/7/21 8:15am to 9:00am</u>		
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Orthodontics

DATE: 15/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
8:00 am	2:00 pm	<i>Sabeena</i>	<i>B.T.</i>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<i>Bacillol</i>	<i>Yes</i>	<i>[Signature]</i>	<i>B.T.</i>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE

4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<i>OP Instruments</i>	<i>CIDEX glutaraldehyde</i>	<i>8:00am / 1:45pm</i>	<i>[Signature]</i>	<i>B.T.</i>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<i>Dragonsha tray</i>	<i>8:15am</i>	<i>9:00am</i>	<i>45min</i>	<i>121°C</i>	<i>30psi</i>	<i>15/7/21 8:15am to 9:00am</i>	<i>[Signature]</i>	<i>B.T.</i>
2	<i>Impression tray</i>	<i>"</i>	<i>"</i>	<i>"</i>	<i>"</i>	<i>"</i>			
3	<i>Orthodontic pliers</i>	<i>"</i>	<i>"</i>	<i>"</i>	<i>"</i>	<i>"</i>	<i>15/7/21 8:15am to 9:00am</i>		
4									
5									
6									
7									



Prof. Dr. Harvey Thomas MDS
Principal
Al Azhar Dental College
Modipurna - 685 695

PREPARED AND APPROVED BY

DEPARTMENT Orthodontics DATE: 16/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
8:00am	2:00pm	Sabaree	B.T.

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
Bacillol	Yes	[Signature]	B.T.

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
Yes	3:00pm	[Signature]	B.T.


4. MANUAL CLEANING OF INSTRUMENTS

Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	OP Instruments	CIDEX glutaraldehyde	8:00am / 1:45pm	[Signature]	B.T.

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	Diagnostic tray	8:15AM	9:00AM	45 min	121°	30 psi	16/7/21 8:15am to 9:00am	[Signature]	B.T.
2	Impression tray	"	"	"	"	"			
3	Orthodontic pliers	"	"	"	"	"	16/7/21 8:15am to 9:00am		
4									
5									
6									
7									



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

PREPARED AND APPROVED BY

DEPARTMENT Orthodontics

DATE: 17/7/21

129

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
9:00 am	2:00 pm	Sabena	B.T.

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
Buclol	Yes	[Signature]	B.T.

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	Op Instrument	Clorox glutaraldehyde	8:00 am / 1:45 pm	[Signature]	B.T.

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	Diagnostic tray	8:15 AM	9:00 AM	45 min	121°C	30 psi	17/7/21 8:15 am to 9:00 am	[Signature]	B.T.
2	Impression tray	"	"	"	"	"			
3	Orthodontic pliers	"	"	"	"	"	17/7/21 9:15 am to 9:30 am		
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MD
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Orthodontics

DATE: 19/7/21

130

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8:00 am</u>	<u>2:00 pm</u>	<u>Sabana</u>	<u>BT</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bucllocl</u>	<u>Yes.</u>	<u>[Signature]</u>	<u>BT</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>Yes</u>	<u>3:00pm</u>	<u>[Signature]</u>	<u>BT</u>

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>op instrument</u>	<u>CODEX glutonolaldehyde</u>	<u>8:00 am / 1:45 pm</u>	<u>[Signature]</u>	<u>BT</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>Diagnostic Instrument</u>	<u>9:15 am</u>	<u>10:00 am</u>	<u>45 min</u>	<u>121°C</u>	<u>30 psi</u>	<u>19/7/21 8:15 am to 9:00 am</u>	<u>[Signature]</u>	<u>BT</u>
2	<u>Impression tray</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>		<u>[Signature]</u>	<u>BT</u>
3	<u>Orthodontic pliers.</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>19/7/21 8:15 am to 9:00 am</u>	<u>[Signature]</u>	<u>BT</u>
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Orthodontics

DATE: 09/7/21

131

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
8:00AM	2:00PM	Sabeen	BT

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
Bicillo?	Yes	[Signature]	BT

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
Basit?	1		

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	O.P. OR's burr/cuts	CIDE* 96% canaldehyde	8:00am / 1:45pm	[Signature]	BT

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	Diagnostic tray	8:15AM	9:00AM	45min	121C	20psi	20/7/21 8:15am to 9:00am	[Signature]	BT
2	Impression tray	"	"	"	"	"			
3	Orthodontic piece	"	"	"	"	"			
4							20/7/21 8:15am to 9:00am	[Signature]	BT
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDC
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Orthodontics

DATE: 21/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8:00 AM</u>	<u>12:00 PM</u>	<u>Sabeena</u>	<u>BT</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bocillol</u>	<u>Yes</u>	<u>[Signature]</u>	<u>BT</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>Yes</u>	<u>3:00pm</u>	<u>[Signature]</u>	<u>BT</u>

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>OP instruments</u>	<u>CIDEX</u> <u>Calcium hydroxide</u>	<u>8:00am</u> <u>1:45pm</u>	<u>[Signature]</u>	<u>BT</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	<u>Diagnostic tray</u>	<u>8:15AM</u>	<u>9:00AM</u>	<u>45min</u>	<u>121C</u>	<u>30psi</u>	<u>21/7/21</u> <u>8:15am to 9:00am</u>	<u>[Signature]</u>	<u>BT</u>
2	<u>Impression tray</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>		<u>[Signature]</u>	<u>BT</u>
3	<u>Orthodontic pliers</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>21/7/21</u> <u>8:15am to 9:00am</u>	<u>[Signature]</u>	<u>BT</u>
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Orthodontics

DATE: 22/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8:00 AM</u>	<u>2:00 PM</u>	<u>Suben</u>	<u>B.T</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillo7</u>	<u>Yes</u>	<u>[Signature]</u>	<u>B.T</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>OP instruments</u>	<u>CIDEX glutaraldehyde.</u>	<u>8:00am / 1:45pm</u>	<u>[Signature]</u>	<u>B.T</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>Diagnostic tray</u>	<u>8:15 AM</u>	<u>9:00 AM</u>	<u>45 min</u>	<u>121°C</u>	<u>30 psi</u>	<u>22/7/21 8:15 am to 9:00 am</u>	<u>[Signature]</u>	
2	<u>Impression tray</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>		<u>[Signature]</u>	<u>B.T</u>
3	<u>Orthodontic pliers</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>22/7/21 8:15 am to 9:00 am</u>	<u>[Signature]</u>	
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Orthodontics

DATE: 23/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
8:00AM	2:00PM	Sabeena.	BT

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
Bacillo 7	Yes	[Signature]	BT

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
Yes	3:00pm.	[Signature]	BT

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	OP instruments	CIDEX glutaraldehyde	8:00am / 1:45pm	[Signature]	BT

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	Diagnosis tray	8:15AM	9:00AM	45min	121°C	30psi	23/7/21 8:15am to 9:00am	[Signature]	BT
2	Impression tray	"	"	"	"	"			
3	Orthodontic piece	"	"	"	"	"	23/7/21 8:15am to 9:00am		
4									
5									
6									
7									

PREPARED AND APPROVED BY



DEPARTMENT Orthodontics

DATE: 24/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8:00 AM</u>	<u>2:00 PM</u>	<u>Saleem</u>	<u>BT</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Betillo 7</u>	<u>Yes</u>	<u>[Signature]</u>	<u>BT</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>OP instruments</u>	<u>CIDEX glutaraldehyde</u>	<u>8:00am / 1:45pm</u>	<u>A</u>	<u>BT</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	<u>Diagnosis tray</u>	<u>8:35AM</u>	<u>9:00AM</u>	<u>45min</u>	<u>121C</u>	<u>35psi</u>	<u>24/7/21 8:15am to 9:00am</u>	<u>[Signature]</u>	<u>BT</u>
2	<u>Impression tray</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>			
3	<u>Orthodontic pieces</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>24/7/21 8:15am to 9:00am</u>		
4		<u>"</u>							
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Orthodontics

DATE: 26/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8:00AM</u>	<u>2:00PM</u>	<u>Sabeen</u>	<u>BT</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Betadine</u>	<u>Yes</u>	<u>[Signature]</u>	<u>BT</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>Yes</u>	<u>9:00pm.</u>		<u>BT</u>

4. MANUAL CLEANING OF INSTRUMENTS


SI. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>OP instruments</u>	<u>CIDEX Glucanolyde</u>	<u>8:00am/1:49pm</u>	<u>[Signature]</u>	<u>BT</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

SI.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>Diagnostic tray</u>	<u>8:15AM</u>	<u>9:00AM</u>	<u>45min</u>	<u>121°C</u>	<u>30psi</u>	<u>26/7/21 8:15am to 9:00am</u>	<u>[Signature]</u>	<u>BT</u>
2	<u>Impression tray</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>			
3	<u>Orthodontic pliers</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>26/7/21 8:15am to 9:00am</u>		
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Orthodontics DATE: 27/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8:00AM</u>	<u>2:00PM</u>	<u>Sabeena</u>	<u>BT</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bocillo</u>	<u>Yes</u>	<u>[Signature]</u>	<u>BT</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE


4. MANUAL CLEANING OF INSTRUMENTS

SI. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>OP instruments</u>	<u>CIDEX glutaraldehyde</u>	<u>8:00am to 1:45pm</u>	<u>[Signature]</u>	<u>BT</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

SI.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>Diagnostic tray</u>	<u>8:15AM</u>	<u>9:00AM</u>	<u>45min</u>	<u>121C</u>	<u>90psi</u>	<u>27/7/21 8:15am to 9:00am</u>	<u>[Signature]</u>	<u>BT</u>
2	<u>Impression tray</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>		<u>[Signature]</u>	<u>BT</u>
3	<u>Orthodontic pliers</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>27/7/21 8:15am to 9:00am</u>	<u>[Signature]</u>	<u>BT</u>
4									
5									
6									
7									



PREPARED AND APPROVED BY

Prof. Dr. Harvey Thomas MDS
Principal
Al Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Ostbodontics

DATE: 20/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8:00 AM</u>	<u>2:00 PM</u>	<u>Suhane</u>	<u>BT</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillo 1</u>	<u>Yes</u>	<u>h</u>	<u>BT</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>Yes</u>	<u>3:00pm.</u>	<u>S</u>	<u>BT</u>

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>OP instrument</u>	<u>CDEX glutaraldehyde</u>	<u>8:00am / 1:45pm</u>	<u>A</u>	<u>BT</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>Diagnostic bag</u>	<u>8:15AM</u>	<u>9:00AM</u>	<u>45min</u>	<u>121°C</u>	<u>30psi</u>	<u>23/7/21 8:15am to 9:00am</u>	<u>[Signature]</u>	
2	<u>Impression bag</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>		<u>[Signature]</u>	<u>BT</u>
3	<u>Ostbodontic pliers</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>23/7/21 8:15am to 9:00am</u>	<u>[Signature]</u>	
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Orthodontics DATE: 30/7/21

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8:00 AM</u>	<u>2:00 PM</u>	<u>Babek</u>	<u>BT</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillo7</u>	<u>Yes</u>	<u>[Signature]</u>	<u>BT</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE
<u>Yes</u>	<u>3:00 pm.</u>	<u>[Signature]</u>	<u>[Signature]</u>

4. MANUAL CLEANING OF INSTRUMENTS


Sl. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>OP instruments</u>	<u>CIDEX Glutaraldehyd.</u>	<u>8:00 am / 1:45 pm</u>	<u>[Signature]</u>	<u>BT</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

Sl.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER.
		START	END	CYCLE LENGTH					
1	<u>Impression tray</u>	<u>8:15 AM</u>	<u>9:00 AM</u>	<u>45 mins</u>	<u>121°c</u>	<u>30 psi</u>	<u>[Stamp]</u>	<u>[Signature]</u>	<u>BT</u>
2	<u>Diagnostic tray</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>			
3	<u>Orthodontic pliers</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>[Stamp]</u>		
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605

DEPARTMENT Orthodontics

DATE: 31/7/21

141

1. FLOOR CLEANING

TIME OF CLEANING		HOUSE KEEPING INCHARGE SIGNATURE	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER
MORNING	NOON		
<u>8:00 AM</u>	<u>2:00 PM</u>	<u>Soblene</u>	<u>B.T</u>

2. CLEANING - DENTAL CHAIR

CHEMICAL /DISINFECTANT USED	CLEANED AFTER EACH PATIENT (YES/NO)	ATTENDER INCHARGE (SIGNATURE)	SIGNATURE OF THE STAFF INCHARGE / HICC MEMBER.
<u>Bacillol</u>	<u>Yes</u>	<u>[Signature]</u>	<u>B.T</u>

3. FUMIGATION

FUMIGATION DONE YES/NO	TIME	SIGNATURE	REMARKS & SIGNATURE OF STAFF IN CHARGE

4. MANUAL CLEANING OF INSTRUMENTS


SI. No.	SET OF INSTRUMENTS	CHEMICAL /DISINFECTANT USED AND IMMERSION TIME	TIME OF THE DAY	ATTENDER INCHARGE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
	<u>OP instruments</u>	<u>CLDEX Gluformaldehyd</u>	<u>8:00am 11:45pm</u>	<u>[Signature]</u>	<u>B.T</u>

PREPARED AND APPROVED BY

5. STERILIZATION LOG SHEET

SI.No	SET OF INSTRUMENTS	TIME			TEMP C/F	PRESSURE	TEMP SENSITIVE INDICATOR	OPERATORS SIGNATURE	SIGNATURE OF THE STAFF INCHARGE/ HICC MEMBER
		START	END	CYCLE LENGTH					
1	<u>Impression tray</u>	<u>8:15 AM</u>	<u>9:00 AM</u>	<u>15 min</u>	<u>121°C</u>	<u>30 psi</u>	<u>31/7/21 8:15 am to 9:00 am</u>	<u>[Signature]</u>	<u>B.T</u>
2	<u>Diagnostic tray</u>	"	"	"	"	"			
3	<u>Orthodontic pliers</u>	"	"	"	"	"	<u>31/7/21 8:15 am to 9:00 am</u>		
4									
5									
6									
7									

PREPARED AND APPROVED BY



Prof. Dr. Harvey Thomas MDS
Principal
Al-Azhar Dental College
Thodupuzha - 685 605